



LEAP[®] Scheme Report EFF036

Effluent, Waste Water, Contaminated Land & Hazardous Waste

January-February 2013

Table 2: Results and z-Scores for Analytes in Test Material 2A

laboratory number	analyte	
	Suspended Solids assigned value: 128 mg/L	
	result, mg/L	z-score
004	125	-0.2
006	104	-1.8
007	128.5	0.1
009	113	-1.1
010		
011	129	0.1
013	133	0.4
016		
018	148.5	1.6
024	120.5	-0.6
027	129	0.1
028	131.5	0.3
030	121	-0.5
032	133	0.4
035	119	-0.7
039	155.50	2.2
041	129	0.1
050	133	0.4
052	127.4	0.0
054	138.6	0.9
060	128.9	0.1
065	124	-0.3
068	91.5	-2.8
071	121.0	-0.5
075	121	-0.5
076	134	0.5
079	125.85	-0.1
081	131	0.3
083	136	0.7

z-scores outside $|z| > 2$ are shown in **bold**, see Section 5