

PROVA INTERLABORATORIO



“Agenti microbiologici nelle acque a diversa destinazione d'uso”

2° CICLO

(MIAD-2)

Anno 2015

RAPPORTO DI PROVA VALUTATIVA INTERLABORATORIO

<p>Coordinatore</p> <p>Dott. Fiorenzo Pastoni</p> 	<p>Responsabile Sezione</p> <p>Dott. Giovanni Perego</p> 
---	---

**Sosp AD-2-A
Escherichia coli MF**

Codice	x	Log(x)	z-score	Segnale
0962	750	2,875	1,07	-
0970	350	2,544	-0,59	-
0978	490	2,690	0,14	-
0979	450	2,653	-0,04	-
0980	258	2,412	-1,25	-
0984	160	2,204	-2,29	W
0991	380	2,580	-0,41	-
0993	54	1,732	-4,65	A
0996	540	2,732	0,35	-
0997	480	2,681	0,10	-
0998	-	-	-	-
0999	600	2,778	0,58	-
1007	-	-	-	-
1010	550	2,740	0,39	-
1013	550	2,740	0,39	-
1015	-	-	-	-
1019	500	2,699	0,19	-
1024	-	-	-	-
1026	650	2,813	0,76	-
1027	-	-	-	-
1035	600	2,778	0,58	-
1041	335	2,525	-0,68	-
1042	-	-	-	-
1043	670	2,826	0,82	-
1045	800	2,903	1,21	-
1056	500	2,699	0,19	-
1057	510	2,708	0,23	-
1058	-	-	-	-
1061	-	-	-	-
1066	10	1,000	-8,31	A
1070	-	-	-	-
1074	-	-	-	-
1077	610	2,785	0,62	-
1079	-	-	-	-
1082	840	2,924	1,31	-
1083	-	-	-	-
1084	-	-	-	-
1088	760	2,881	1,09	-
1089	280	2,447	-1,07	-
1090	-	-	-	-
1095	440	2,643	-0,09	-
1097	375	2,574	-0,44	-
1099	290	2,462	-1,00	-
1101	510	2,708	0,23	-
1102	-	-	-	-
1104	360	2,556	-0,53	-