



LEAP[®] Scheme Report MICRO61

Potable Water

October 2012

NOT CONTROLLED WHEN PRINTED

Table 3: Results and z-Scores for Colony Count after 2 days at 37°C

laboratory number	examination	
	colony count assigned value 197 cfu/mL	
	result cfu/mL	z-score
008	190	-0.1
011	170	-0.6
014	283	2.0
018	165	-0.7
019	225	0.6
020	202	0.1
021	97	-2.3
028	228	0.7
032	198 +/-20	0.0
033	201 +/-20	0.1
035	139	-1.3
036	238	0.9
037	197	0.0
038	223	0.6
039	200	0.1
041	158	-0.9
042	>300	
043	260	1.4
044	150	-1.1

z-scores outside $|z| > 2$ are shown in **bold**, see Section 5